

MS#1

Evaluation of intubation using the Airtraq or Macintosh laryngoscope by anaesthetists in easy and simulated difficult laryngoscopy – a manikin study*

Anaesthesia, 2006; 51:469-477.

Maharaj CH et Al. Univ. Of Ireland, Galway

In the simulated difficult laryngoscopy scenarios, the AIRTRAQ was **more successful** in achieving tracheal intubation, required **less time to intubate** successfully, caused **less dental trauma**, and was considered by the anaesthetists to be **easier to use**.

